



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 7, 2011 INITIALS: Dev OWNER: CITY OF BELLEVUE

DISTRICT: ADDRESS:

LOCATION: COUNTY:

MAKE: GEC TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII

MODEL: MD-500 ID. NO: 4355 CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIAL PLATFORM 10"X11"

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| CENTER        | 5000        | 5000             | 0     |               |             |                  |       |
| CENTER        | 10000       | 10000            | 0     |               |             |                  |       |
| CENTER        | 15000       | 15000            | 0     |               |             |                  |       |
| CENTER        | 20000       | 19900            | -100  |               |             |                  |       |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration", Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Mar 21, 2011

Print name: DAVID CROMER

State of Washington  
County of Shuster

Signed and sworn to (or affirmed before  
me on Mar 21, 2011 by

Brenda A Biscay  
(Signature of notary public)

Brenda A. Biscay

Printed Name of Notary

My appointment expires 3-12-12

